

# Accuracy Certification

Certification NO.: 202405000030

Calibration Date: 05/29/2024

## Calibration Instructions:

The above products have thoroughly comprehensive and detailed testing and verification in the laboratory. All reference instruments are calibrated by Guangdong 602 Measurement and Testing CO.,LTD, a CNAS (China National Accreditation Service) accredited laboratory, and traceable to NIST according Mutual Recognition Agreement.

Device	Model	Accuracy
Comparative testing equipment	HP32	±0.5 %RH, ±0.1 °C
Constant Temperature and Humidity Equipment	HT-S150M	±0.1 %RH, ±0.1 °C
Constant temperature freezer equipment	HT-DW-28	±0.1 °C

Model No.	Serial No	Production Date	Expiry Date	Quantity
Innolog	202405006051-202405006100	05/2024	05/2025	50pcs

Temperature Accuracy -20 °C ~ 40 °C (±0.5 °C)      Others ±1

Validation	Indicated	Variance	Tolerance	Result
-30 °C	-30.3 °C	-0.3 °C	±1 °C	PASS
-20 °C	-19.7 °C	0.3 °C	±0.5 °C	PASS
-10 °C	-10.2 °C	-0.2 °C	±0.5 °C	PASS
0 °C	0.2 °C	0.2 °C	±0.5 °C	PASS
10 °C	10.1 °C	0.1 °C	±0.5 °C	PASS
20 °C	19.9 °C	-0.1 °C	±0.5 °C	PASS
40 °C	40.3 °C	0.3 °C	±0.5 °C	PASS
60 °C	60.7 °C	0.7 °C	±1 °C	PASS
70 °C	70.5 °C	0.5 °C	±1 °C	PASS

Kelvin Solutions

INSPECTED BY *Andy*